



## RNA/small RNA sample preparation guidelines

### RNA for RNA-Seq / smallRNA-Seq application

IGATech offers nucleic acids extraction service and we can set up a dedicated extraction workflow for your specific substrate. Please enquire.

#### General indications for RNA submission

In order to obtain a high quality sequencing data, customers must provide a good quality RNA, in detail:

- the 260/280 ratio of your RNA sample should be >1.8;
- samples should be resuspended in nuclease-free water;
- high-quality RNA should have two prominent bands (e.g. ribosomal RNA) on agarose gel; intensity of the 28S band (at 4.5 kb) should be twice of the 18S one (at 1.9 kb);
- RNA should have an RNA Integrity Number (RIN) > 8 on an Agilent Bioanalyzer 2100.
- RNA samples must be free of contaminating proteins and other cellular material, organic solvents (including phenol and ethanol) and salts. We strongly recommend commercial kits for totalRNA or miRNA extraction (e.g. Spectrum Plant Total RNA Kit, TRI-REAGENT, RNeasy, MirVANA or MirPremier).

Please note that use of degraded RNA can result in low yield, over-representation of 3' ends of the RNA molecules or failure of the protocol.

Customers need to provide the result analysis of Agilent 2100 Bioanalyzer or, at least, gel-electrophoresis image showing the RNA quality. Quantify your RNA samples by spectrophotometer (e.g. Nanodrop) or fluorimeter (e.g. QuBit).

Sample acceptance is based on the amount, concentration and Bioanalyzer QC measured by us. Send RNA samples in 1.5-2ml Eppendorf tubes sealed with parafilm (0.2 mL and 0.5mL tubes as well as strip tubes WILL BE NOT ACCEPTED). If you have 24 samples or more, please put your samples in a 96-wells skirted plate (MANDATORY). Do not ship plates without secondary containment as these may crack when placed directly on dry ice.



The tubes must have, on the vial top, a clear and permanent sign (or a thin label) with the progressive number of the mailed samples according to the LABEL\_ID in the **Samples Spread Sheet**. Send RNA samples in dry-ice or lyophilized with RNastable (Biomatrica).

### Protocol-specific requirements

#### **RNA for TruSeq Stranded mRNA-seq**

The total amount requested is 2.5 µg (minimum concentration of 40 ng/µL, minimum volume of 20 µL). For *de novo* application, we suggest to send a minimum of 4 µg.

#### **RNA for TruSeq Stranded Total RNA-Seq**

The total amount requested is 1 µg (minimum concentration of 50 ng/µL, minimum volume of 20 µL). A DNase I step is mandatory after the RNA isolation. RNA that has DNA contamination will result in an underestimation of the amount of the RNA used and poor data quality. For *de novo* application, we suggest to send a minimum of 2 µg.

For Total RNA applications, we usually perform depletion of ribosomal RNA with one of the following Illumina modules according to the species (information can be also found [here](#)):

- Ribo-Zero Gold Human/Mouse/Rat
- Ribo-Zero Plant system
- Ribo-Zero Gram-Positive Bacteria
- Ribo-Zero Gram-NegativeBacteria
- Ribo-Zero Yeast

For any modification of our standard workflow please inquire.

For FFPE samples we accept down to a total amount of 200 ng (minimum concentration of 10 ng/µl, minimum volume of 20 µL). Total RNA protocol also works with degraded RNAs and FFPE RNAs even if the success rate is not guarantee.



### **RNA for TruSeq smallRNA-Seq**

The total amount requested is 2 µg (minimum concentration of 200 ng/µl) of total RNA or 100 ng of previously isolated microRNA (minimum concentration of 10 ng/µl) in 10 µl of nuclease-free water or 10 mM Tris-HCl, pH 8.5.

### **RNA for Nugen SoLo RNA-Seq System**

The total amount requested is 200 pg in at least 20 uL low-EDTA TE buffer (recommended) or nuclease-free water. We can use (experimentally) quantities below 100pg but the rate of duplicates is going to be above 50%, based on our tests. Please inquire for custom designs for depletion of unwanted transcripts (rRNA, abundant constitutive mRNA) with InDA-C system.

Please, do not forget to send us compiled **Samples Spreadsheet** and the **Customer registration form**, together with shipped samples as well as via e-mail.